



## Connection Diagram of Iotron™ Sensors without Preamplifiers to Jenco 6311 pH / ORP Transmitter & Controller

**Connection from Iotron™ Sensor to Terminal Block on back of Transmitter (READ CAREFULLY)**

<u>Cable Color Coding</u>	<u>Sensor Cable Lead Value</u>	<u>Terminal Label As given on Meter</u>	<u>Terminal Value (See Diagram Below)</u>
Black	TC Input	ATC (-)	ATC IN (No Polarity)
Black	TC Input	ATC (+)	ATC IN (No Polarity)
Red	Reference Input	pH (-)	Reference
Clear	pH/ORP Input (Signal)	pH (+)	Glass



Note 1: Temperature Compensator is either a 3000 Ohm Balco or 1000 Ohm Platinum Element. You must set this TC Value in the Calibration Mode (if not preconfigured). Use Code: **rES** for the AtC setting if a 3000 Ohm Balco Resistor is present or use Code: **Pt** for the AtC setting if a 1000 Ohm Platinum Element is present

Note 2: The Temperature Compensator element employed can be determined by measuring the resistance between the two black wires on any multimeter (in Ohms).

Note 3: Your first pH buffer value must be 7.00 or 6.86! The second pH buffer value can be 4.01 or 10.00 if starting with pH buffer 7.00 as the pH buffer 1. The second pH buffer value can be 4.01 and 9.18 if starting with pH buffer 6.86 as the pH buffer 1. Consult the operation manual for further calibration and power hook-up details.